

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 00757 **Calib. Date:** 09/16/2011 **Calib. No.:** 00017
Certification File No.: 00758 **Cert. Date:** 09/16/2011 **Cert. No.:** 00013
Linearity File No.: 00759 **Lin. Date:** 09/16/2011 **Lin. No.:** 00013
Solution File No.: 00767 **Soln. Date:** 01/09/2012 **Soln. No.:** 00123
Sequential File No.: 00767 **File Date:** 01/09/2012

Calibrating Unit: WET **Model No.:** CU-34 **Serial No.:** DDUL S3-0183
Control Solution %: 0.100% **Expires:** 07/07/2013
Solution Control Lot: 11G093 **Bottle No.:** 0884

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:54S		
Control 1 EC	0.100%	16:55S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:55S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:56S		
Control 2 EC	0.099%	16:56S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:56S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:57S		
Control 3 EC	0.100%	16:58S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:58S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:59S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: 

Badge No.: 1120

Date: 01/09/2012