

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	00781	Calib. Date:	03/01/2012
Certification File No.:	00782	Cert. No.:	00018
Linearity File No.:	00783	Cert. Date:	03/01/2012
Solution File No.:	00791	Lin. Date:	03/01/2012
Sequential File No.:	00791	Lin. No.:	00014
		Soln. Date:	03/21/2012
		Soln. No.:	00127
		File Date:	03/21/2012

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	07/07/2013
Solution Control Lot:	11G093			Bottle No.:	0890

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:02S		
Control 1 EC	0.100%	15:02S	34.1°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:02S	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03S		
Control 2 EC	0.099%	15:04S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	15:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:04S		
Control 3 EC	0.099%	15:05S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	15:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:06S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: 

Badge No.: 1120

Date: 03/21/2012