

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00781	Calib. Date: 03/01/2012
Certification File No.:	00782	Calib. No.: 00018
Linearity File No.:	00783	Cert. Date: 03/01/2012
Solution File No.:	00832	Cert. No.: 00014
Sequential File No.:	00832	Lin. Date: 03/01/2012
		Lin. No.: 00014
		Soln. Date: 08/16/2012
		Soln. No.: 00133
		File Date: 08/16/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	11G093	Expires: 07/07/2013
		Bottle No.: 0895

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:40D		
Control 1 EC	0.102%	02:41D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	02:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:42D		
Control 2 EC	0.101%	02:42D	34.1°C	*** TEST PASSED ***
Control 2 IR	0.102%	02:42D	34.1°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:43D		
Control 3 EC	0.100%	02:44D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	02:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:44D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: NIRO

First Name: SGT

MI: J

Signature: \_\_\_\_\_



Badge No.: 1114

Date: 08/16/2012