

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 00835 Calib. Date: 08/31/2012 Calib. No.: 00019
Certification File No.: 00836 Cert. Date: 08/31/2012 Cert. No.: 00015
Linearity File No.: 00837 Lin. Date: 08/31/2012 Lin. No.: 00015
Solution File No.: 00843 Soln. Date: 09/27/2012 Soln. No.: 00136
Sequential File No.: 00843 File Date: 09/27/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 07/07/2013
Solution Control Lot: 11G093 Bottle No.: 0877

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:10D		
Control 1 EC	0.097%	11:10D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	11:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:13D		
Control 2 EC	0.096%	11:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	11:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:14D		
Control 3 EC	0.097%	11:15D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: 

Badge No.: 1120

Date: 09/27/2012