

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 00835 Calib. Date: 08/31/2012 Calib. No.: 00019  
Certification File No.: 00836 Cert. Date: 08/31/2012 Cert. No.: 00015  
Linearity File No.: 00837 Lin. Date: 08/31/2012 Lin. No.: 00015  
Solution File No.: 00847 Soln. Date: 11/01/2012 Soln. No.: 00137  
Sequential File No.: 00847 File Date: 11/01/2012

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183  
Control Solution %: 0.100% Expires: 07/07/2013  
Solution Control Lot: 11G093 Bottle No.: 0881

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:15S		
Control 1 EC	0.100%	00:15S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	00:15S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:16S		
Control 2 EC	0.098%	00:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:17S		
Control 3 EC	0.099%	00:18S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	00:18S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:19S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: 

Badge No.: 1120

Date: 11/01/2012