

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00835	Calib. Date: 08/31/2012
Certification File No.:	00836	Calib. No.: 00019
Linearity File No.:	00837	Cert. Date: 08/31/2012
Solution File No.:	00855	Cert. No.: 00015
Sequential File No.:	00855	Lin. Date: 08/31/2012
		Lin. No.: 00015
		Soln. Date: 11/19/2012
		Soln. No.: 00138
		File Date: 11/19/2012
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	11G093	Expires: 07/07/2013
		Bottle No.: 0882

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:46S		
Control 1 EC	0.099%	15:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	15:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:48S		
Control 2 EC	0.098%	15:48S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:49S		
Control 3 EC	0.098%	15:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	15:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:51S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: 

Badge No.: 1120

Date: 11/19/2012