

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE		
Calibration File No.:	00835	Calib. Date:	08/31/2012
Certification File No.:	00836	Cert. Date:	08/31/2012
Linearity File No.:	00837	Lin. Date:	08/31/2012
Solution File No.:	00869	Soln. Date:	12/15/2012
Sequential File No.:	00869	File Date:	12/15/2012
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUL S3-0183
Solution Control Lot:	11G093	Expires:	07/07/2013
		Bottle No.:	0892

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:28S		
Control 1 EC	0.101%	04:28S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:29S		
Control 2 EC	0.099%	04:30S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	04:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:30S		
Control 3 EC	0.099%	04:31S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:31S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: RUNGE

First Name: GLEN

MI: L

Signature: _____



Badge No.: 1118

Date: 12/15/2012