Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARTN-0042
Location:	HIGHTSTOWN POLICE				
Calibration File No.:	00994	Calib. Date:	03/13/2014	Calib. No.:	00024
Certification File No.:	00995	Cert. Date:	03/13/2014	Cert. No.:	00019
Linearity File No.:	00996	Lin. Date:	03/13/2014	Lin. No.:	00019
Solution File No.:	00997	Soln. Date:	03/13/2014	Soln. No.:	00165
Sequential File No.:	00997	File Date:	03/13/2014		
Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0182
Control Solution %:	0.100%			Expires:	09/13/2014
Solution Control Lot:	12I107			Bottle No.:	0856

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:46S 10:46D MG		
Control 1 EC	0.100%	09:47S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	09:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:48S		
Control 2 EC	0.099%	09:49S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:49S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:50S		
Control 3 EC	0.098%	09:50S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	09:50S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:51S 10:51D MG		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJP2-035 MG

Changed By:

Last Name: GONCALVES First Name: MICHELLE MI: L

Signature: A/Sgt. Gonc A6040 Badge No.: 6040

Date: 03/13/2014