

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01070	Calib. Date: 09/02/2014
Certification File No.:	01071	Calib. No.: 00025
Linearity File No.:	01072	Cert. Date: 09/02/2014
Solution File No.:	01088	Cert. No.: 00020
Sequential File No.:	01088	Lin. Date: 09/02/2014
		Lin. No.: 00020
		Soln. Date: 12/04/2014
		Soln. No.: 00177
		File Date: 12/04/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	14F130	Expires: 06/20/2016
		Bottle No.: 0069

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:45S		
Control 1 EC	-.---%	16:46S	22.5°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	16:46S	22.5°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: \_\_\_\_\_

Badge No.: 1113

Date: 12/04/2014

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01070 Calib. Date: 09/02/2014 Calib. No.: 00025  
Certification File No.: 01071 Cert. Date: 09/02/2014 Cert. No.: 00020  
Linearity File No.: 01072 Lin. Date: 09/02/2014 Lin. No.: 00020  
Solution File No.: 01089 Soln. Date: 12/04/2014 Soln. No.: 00178  
Sequential File No.: 01089 File Date: 12/04/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183  
Control Solution %: 0.100% Expires: 06/20/2016  
Solution Control Lot: 14F130 Bottle No.: 0075

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:40S		
Control 1 EC	0.103%	18:41S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.103%	18:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:41S		
Control 2 EC	0.102%	18:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	18:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:43S		
Control 3 EC	0.102%	18:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.103%	18:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:44S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: 

Badge No.: 1113

Date: 12/04/2014