

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01070 Calib. Date: 09/02/2014 Calib. No.: 00025
Certification File No.: 01071 Cert. Date: 09/02/2014 Cert. No.: 00020
Linearity File No.: 01072 Lin. Date: 09/02/2014 Lin. No.: 00020
Solution File No.: 01078 Soln. Date: 10/03/2014 Soln. No.: 00175
Sequential File No.: 01078 File Date: 10/03/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 1168

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:54D		
Control 1 EC	0.101%	13:55D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.102%	13:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:55D		
Control 2 EC	0.101%	13:56D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	13:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:57D		
Control 3 EC	0.101%	13:57D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	13:57D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:58D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: Sgt. J. Miller #1113

Badge No.: 1113

Date: 10/03/2014