

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 00994 Calib. Date: 03/13/2014 Calib. No.: 00024  
Certification File No.: 00995 Cert. Date: 03/13/2014 Cert. No.: 00019  
Linearity File No.: 00996 Lin. Date: 03/13/2014 Lin. No.: 00019  
Solution File No.: 01021 Soln. Date: 04/23/2014 Soln. No.: 00168  
Sequential File No.: 01021 File Date: 04/23/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183  
Control Solution %: 0.100% Expires: 09/13/2014  
Solution Control Lot: 12I107 Bottle No.: 0852

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:04D		
Control 1 EC	0.097%	14:04D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:04D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:05D		
Control 2 EC	0.097%	14:06D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:06D		
Control 3 EC	0.097%	14:07D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:08D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GENDRON

First Name: FRANK

MI: G

Signature: 

Badge No.: 1120

Date: 04/23/2014