

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00994	Calib. Date: 03/13/2014
Certification File No.:	00995	Calib. No.: 00024
Linearity File No.:	00996	Cert. Date: 03/13/2014
Solution File No.:	01033	Cert. No.: 00019
Sequential File No.:	01033	Lin. Date: 03/13/2014
		Lin. No.: 00019
		Soln. Date: 05/17/2014
		Soln. No.: 00169
		File Date: 05/17/2014

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUL S3-0183
Control Solution %:	0.100%			Expires:	09/13/2014
Solution Control Lot:	12I107			Bottle No.:	0855

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:41D		
Control 1 EC	0.097%	07:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	07:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:43D		
Control 2 EC	0.096%	07:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	07:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:44D		
Control 3 EC	0.097%	07:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	07:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:45D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: _____

Sgt. Benjamin Miller #1113

Badge No.: 1113

Date: 05/17/2014