

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 00994 Calib. Date: 03/13/2014 Calib. No.: 00024
Certification File No.: 00995 Cert. Date: 03/13/2014 Cert. No.: 00019
Linearity File No.: 00996 Lin. Date: 03/13/2014 Lin. No.: 00019
Solution File No.: 01062 Soln. Date: 08/17/2014 Soln. No.: 00173
Sequential File No.: 01062 File Date: 08/17/2014

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 11/05/2014
Solution Control Lot: 12K108 Bottle No.: 1173

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:29D		
Control 1 EC	0.097%	19:30D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.097%	19:30D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:30D		
Control 2 EC	0.096%	19:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	19:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:32D		
Control 3 EC	0.096%	19:32D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.097%	19:32D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:33D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: OFC

MI: R

Signature: _____



Badge No.: 1123

Date: 08/17/2014