

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	00994	Calib. Date: 03/13/2014
Certification File No.:	00995	Cert. No.: 00024
Linearity File No.:	00996	Cert. Date: 03/13/2014
Solution File No.:	01009	Lin. No.: 00019
Sequential File No.:	01009	Soln. Date: 04/03/2014
		Soln. No.: 00167
		File Date: 04/03/2014
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	12I107	Expires: 09/13/2014
		Bottle No.: 0860

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:33S	14:33 D MG	
Control 1 EC	0.098%	13:34S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	13:34S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:34S		
Control 2 EC	0.098%	13:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	13:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:36S		
Control 3 EC	0.098%	13:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	13:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:37S	14:37 D MG	

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJP2-055 (MG)

Changed By:

Last Name: GONCALVES


First Name: MICHELLE

MI: L

Signature: A/Sgt. Michel #6040

Badge No.: 6040

Date: 04/03/2014



Dräger

CERTIFICATE OF ACCURACY

This Certificate of Accuracy verifies that the specified unit has been examined and found to be in compliance with National Highway and Traffic Safety Administration regulations for devices used to calibrate Evidential Breath Testers.
(F.R. Vol. 59 No. 249 12/19/94 Notices)

Draeger Safety Diagnostics, Inc.

- Model: ALCOTEST® CU34
- Model: MARK IIA
- Other: _____

Serial Number:

DDUL 53-0183

Certification Date

3-24-14

Technician

ROC

Re-Certification Due Date

3-24-15

Dräger

ALCOTEST® 7110 TEMPERATURE PROBE

CERTIFICATE OF ACCURACY

This is to certify that the Alcotest® 7110 Temperature Probe has been tested for accuracy with instrumentation that is traceable to the National Institute of Standards and Technology (NIST).

The manufacturer recommends accuracy verification of the Temperature Probe, within 12 months of the certification date below, or sooner, according to your State Specification.

For accurate temperature readings, the probe value on this certificate, noted below, must be programmed into the Alcotest® 7110.

Serial Number Temp. Probe

DDWJF2-055

Certification date:

3-24-14

Next Certification due:

3-24-15

Probe Value

102

Draeger Safety Diagnostics, Inc.
Technical Service Department

ROC