

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01240	Calib. Date: 10/26/2016
Certification File No.:	01241	Calib. No.: 00030
Linearity File No.:	01242	Cert. Date: 10/26/2016
Solution File No.:	01253	Cert. No.: 00025
Sequential File No.:	01253	Lin. Date: 10/26/2016
		Lin. No.: 00025
		Soln. Date: 02/26/2017
		Soln. No.: 00215
		File Date: 02/26/2017

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0182
Control Solution %:	0.100%		Expires: 05/17/2018
Solution Control Lot:	201605D		Bottle No.: 0792

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:25S		
Control 1 EC	0.101%	22:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:27S		
Control 2 EC	0.099%	22:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:28S		
Control 3 EC	0.100%	22:29S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:29S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: _____



Badge No.: 1110

Date: 02/26/2017