

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01240 Calib. Date: 10/26/2016 Calib. No.: 00030  
Certification File No.: 01241 Cert. Date: 10/26/2016 Cert. No.: 00025  
Linearity File No.: 01242 Lin. Date: 10/26/2016 Lin. No.: 00025  
Solution File No.: 01248 Soln. Date: 01/28/2017 Soln. No.: 00214  
Sequential File No.: 01248 File Date: 01/28/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182  
Control Solution %: 0.100% Expires: 05/17/2018  
Solution Control Lot: 201605D Bottle No.: 0794

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:03S		
Control 1 EC	0.102%	07:04S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:04S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:05S		
Control 2 EC	0.100%	07:06S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	07:06S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:06S		
Control 3 EC	0.100%	07:07S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	07:07S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


## Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: \_\_\_\_\_



Badge No.: 1113

Date: 01/28/2017