

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01284 Calib. Date: 08/07/2017 Calib. No.: 00032
Certification File No.: 01285 Cert. Date: 08/07/2017 Cert. No.: 00027
Linearity File No.: 01286 Lin. Date: 08/07/2017 Lin. No.: 00027
Solution File No.: 01303 Soln. Date: 10/25/2017 Soln. No.: 00225
Sequential File No.: 01303 File Date: 10/25/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 0922

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:06D		
Control 1 EC	0.098%	12:07D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	12:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07D		
Control 2 EC	0.097%	12:08D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	12:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:09D		
Control 3 EC	0.098%	12:10D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:10D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: _____

SJA Miller #1113

Badge No.: 1113

Date: 10/25/2017