

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01284	Calib. Date: 08/07/2017
Certification File No.:	01285	Calib. No.: 00032
Linearity File No.:	01286	Cert. Date: 08/07/2017
Solution File No.:	01306	Cert. No.: 00027
Sequential File No.:	01306	Lin. Date: 08/07/2017
		Lin. No.: 00027
		Soln. Date: 11/24/2017
		Soln. No.: 00226
		File Date: 11/24/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1065

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:26S		
Control 1 EC	-.---%	18:27S	19.8°C	SIMULATOR TEMP. ERROR
Control 1 IR	-.---%	18:27S	19.8°C	SIMULATOR TEMP. ERROR
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 11/24/2017

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01284	Calib. Date: 08/07/2017
Certification File No.:	01285	Calib. No.: 00032
Linearity File No.:	01286	Cert. Date: 08/07/2017
Solution File No.:	01307	Cert. No.: 00027
Sequential File No.:	01307	Lin. Date: 08/07/2017
		Lin. No.: 00027
		Soln. Date: 11/24/2017
		Soln. No.: 00227
		File Date: 11/24/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 0925

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:22S		
Control 1 EC	0.098%	21:23S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:24S		
Control 2 EC	0.097%	21:25S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:25S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:25S		
Control 3 EC	0.097%	21:26S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:27S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 11/24/2017