

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01284	Calib. Date: 08/07/2017
Certification File No.:	01285	Calib. No.: 00032
Linearity File No.:	01286	Cert. Date: 08/07/2017
Solution File No.:	01313	Cert. No.: 00027
Sequential File No.:	01313	Lin. Date: 08/07/2017
		Lin. No.: 00027
		Soln. Date: 12/17/2017
		Soln. No.: 00228
		File Date: 12/17/2017
Calibrating Unit:	WET	Módel No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 0909

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	05:49S		
Control 1 EC	-.---%	05:49S	33.8°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	05:49S	33.8°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: \_\_\_\_\_

Badge No.: 1110

Date: 12/17/2017

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01284 Calib. Date: 08/07/2017 Calib. No.: 00032  
Certification File No.: 01285 Cert. Date: 08/07/2017 Cert. No.: 00027  
Linearity File No.: 01286 Lin. Date: 08/07/2017 Lin. No.: 00027  
Solution File No.: 01314 Soln. Date: 12/17/2017 Soln. No.: 00229  
Sequential File No.: 01314 File Date: 12/17/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183  
Control Solution %: 0.100% Expires: 09/14/2019  
Solution Control Lot: 17290 Bottle No.: 0789

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:10S		
Control 1 EC	0.098%	21:11S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:11S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:12S		
Control 2 EC	0.096%	21:12S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:12S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:13S		
Control 3 EC	0.096%	21:14S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.097%	21:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:15S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


### Changed By:

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature:



Badge No.: 1110

Date: 12/17/2017