

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01254	Calib. Date: 03/03/2017
Certification File No.:	01255	Calib. No.: 00031
Linearity File No.:	01256	Cert. Date: 03/03/2017
Solution File No.:	01260	Cert. No.: 00026
Sequential File No.:	01260	Lin. Date: 03/03/2017
		Lin. No.: 00026
		Soln. Date: 03/31/2017
		Soln. No.: 00217
		File Date: 03/31/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 0988

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:18S		
Control 1 EC	0.099%	19:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:19S		
Control 2 EC	0.097%	19:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	19:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:21S		
Control 3 EC	0.097%	19:22S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	19:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:23S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 03/31/2017