

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01254 Calib. Date: 03/03/2017 Calib. No.: 00031
Certification File No.: 01255 Cert. Date: 03/03/2017 Cert. No.: 00026
Linearity File No.: 01256 Lin. Date: 03/03/2017 Lin. No.: 00026
Solution File No.: 01262 Soln. Date: 04/23/2017 Soln. No.: 00218
Sequential File No.: 01262 File Date: 04/23/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182
Control Solution %: 0.100% Expires: 01/30/2019
Solution Control Lot: 17050 Bottle No.: 0960

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:42D		
Control 1 EC	0.102%	10:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:43D		
Control 2 EC	0.099%	10:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:45D		
Control 3 EC	0.100%	10:46D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	10:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____



Badge No.: 1123

Date: 04/23/2017