

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01284	Calib. Date: 08/07/2017
Certification File No.:	01285	Calib. No.: 00032
Linearity File No.:	01286	Cert. Date: 08/07/2017
Solution File No.:	01295	Cert. No.: 00027
Sequential File No.:	01295	Lin. Date: 08/07/2017
		Lin. No.: 00027
		Soln. Date: 08/28/2017
		Soln. No.: 00223
		File Date: 08/28/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	17050	Expires: 01/30/2019
		Bottle No.: 1037

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:52D		
Control 1 EC	0.098%	00:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:54D		
Control 2 EC	0.097%	00:55D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	00:55D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:55D		
Control 3 EC	0.098%	00:56D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	00:56D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: JIMENEZ

First Name: SGT.

MI: F.

Signature: \_\_\_\_\_

SGT. F. Jimenez #1110

Badge No.: 1110

Date: 08/28/2017