

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01477	Calib. Date: 10/04/2019
Certification File No.:	01478	Calib. No.: 00037
Linearity File No.:	01479	Cert. Date: 10/04/2019
Solution File No.:	01488	Cert. No.: 00032
Sequential File No.:	01488	Lin. Date: 10/04/2019
		Lin. No.: 00032
		Soln. Date: 10/31/2019
		Soln. No.: 00260
		File Date: 10/31/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0565

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	01:01S		
Control 1 EC	0.100%	01:02S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	01:02S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:02S		
Control 2 EC	0.099%	01:03S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	01:03S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:04S		
Control 3 EC	0.100%	01:05S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	01:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	01:05S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_

Badge No.: 1123

Date: 10/31/2019