

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01405 Calib. Date: 11/28/2018 Calib. No.: 00035
Certification File No.: 01406 Cert. Date: 11/28/2018 Cert. No.: 00030
Linearity File No.: 01407 Lin. Date: 11/28/2018 Lin. No.: 00030
Solution File No.: 01420 Soln. Date: 02/16/2019 Soln. No.: 00249
Sequential File No.: 01420 File Date: 02/16/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0183
Control Solution %: 0.100% Expires: 01/31/2020
Solution Control Lot: 18050 Bottle No.: 0896

| Function | Result | Time | Temperature | Comment(s) |
|-------------------|--------|--------|----------------|---------------------|
| | %BAC | HH:MM | Simulator (°C) | or Error(s) |
| Ambient Air Blank | 0.000% | 04:40S | | |
| Control 1 EC | 0.100% | 04:40S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 04:40S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:42S | | |
| Control 2 EC | 0.097% | 04:42S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 04:42S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:43S | | |
| Control 3 EC | 0.098% | 04:44S | 33.9°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 04:44S | 33.9°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 04:45S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 02/16/2019