

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01405	Calib. Date: 11/28/2018
Certification File No.:	01406	Calib. No.: 00035
Linearity File No.:	01407	Cert. Date: 11/28/2018
Solution File No.:	01422	Cert. No.: 00030
Sequential File No.:	01422	Lin. Date: 11/28/2018
		Lin. No.: 00030
		Soln. Date: 03/12/2019
		Soln. No.: 00250
		File Date: 03/12/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	18050	Expires: 01/31/2020
		Bottle No.: 0889

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:18S		
Control 1 EC	0.101%	18:19S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:20S		
Control 2 EC	0.097%	18:21S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	18:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:22S		
Control 3 EC	0.099%	18:23S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	18:23S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:24S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 03/12/2019