

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01433	Calib. Date: 05/16/2019
Certification File No.:	01434	Calib. No.: 00036
Linearity File No.:	01435	Cert. Date: 05/16/2019
Solution File No.:	01447	Cert. No.: 00031
Sequential File No.:	01447	Lin. Date: 05/16/2019
		Lin. No.: 00031
		Soln. Date: 06/13/2019
		Soln. No.: 00254
		File Date: 06/13/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0506

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:18D		
Control 1 EC	0.100%	06:19D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	06:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:20D		
Control 2 EC	0.099%	06:20D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:20D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:21D		
Control 3 EC	0.099%	06:22D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	06:22D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:23D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_

Badge No.: 1123

Date: 06/13/2019