

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01433 Calib. Date: 05/16/2019 Calib. No.: 00036  
Certification File No.: 01434 Cert. Date: 05/16/2019 Cert. No.: 00031  
Linearity File No.: 01435 Lin. Date: 05/16/2019 Lin. No.: 00031  
Solution File No.: 01454 Soln. Date: 07/08/2019 Soln. No.: 00255  
Sequential File No.: 01454 File Date: 07/08/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182  
Control Solution %: 0.100% Expires: 01/08/2021  
Solution Control Lot: 19020 Bottle No.: 0629

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:50D		
Control 1 EC	0.100%	04:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	04:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:52D		
Control 2 EC	0.099%	04:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:53D		
Control 3 EC	0.099%	04:54D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	04:54D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:54D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: 

Badge No.: 1123

Date: 07/08/2019