

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01433	Calib. Date: 05/16/2019
Certification File No.:	01434	Calib. No.: 00036
Linearity File No.:	01435	Cert. Date: 05/16/2019
Solution File No.:	01475	Cert. No.: 00031
Sequential File No.:	01475	Lin. Date: 05/16/2019
		Lin. No.: 00031
		Soln. Date: 09/23/2019
		Soln. No.: 00258
		File Date: 09/23/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	19020	Expires: 01/08/2021
		Bottle No.: 0498

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	07:09D		
Control 1 EC	0.099%	07:09D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	07:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:10D		
Control 2 EC	0.098%	07:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	07:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:11D		
Control 3 EC	0.098%	07:12D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	07:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	07:13D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: _____



Badge No.: 1113

Date: 09/23/2019