

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARTN-0042
Location: HIGHTSTOWN POLICE
Calibration File No.: 01701 Calib. Date: 06/06/2023 Calib. No.: 00047
Certification File No.: 01702 Cert. Date: 06/06/2023 Cert. No.: 00042
Linearity File No.: 01703 Lin. Date: 06/06/2023 Lin. No.: 00042
Solution File No.: 01715 Soln. Date: 08/04/2023 Soln. No.: 00320
Sequential File No.: 01715 File Date: 08/04/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182
Control Solution %: 0.100% Expires: 04/12/2025
Solution Control Lot: 23110 Bottle No.: 1216

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:38D		
Control 1 EC	0.101%	10:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	10:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:39D		
Control 2 EC	0.101%	10:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	10:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:41D		
Control 3 EC	0.101%	10:42D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	10:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature:  #1113

Badge No.: 1113

Date: 08/04/2023