

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01673 Calib. Date: 09/02/2022 Calib. No.: 00045  
Certification File No.: 01674 Cert. Date: 09/02/2022 Cert. No.: 00040  
Linearity File No.: 01675 Lin. Date: 09/02/2022 Lin. No.: 00040  
Solution File No.: 01685 Soln. Date: 01/03/2023 Soln. No.: 00312  
Sequential File No.: 01685 File Date: 01/03/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182  
Control Solution %: 0.100% Expires: 05/09/2024  
Solution Control Lot: 22160 Bottle No.: 1436

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:32S		
Control 1 EC	0.100%	12:33S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:34S		
Control 2 EC	0.097%	12:35S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	12:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:35S		
Control 3 EC	0.098%	12:36S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:36S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:37S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature:  #1113

Badge No.: 1113

Date: 01/03/2023