

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01733	Calib. Date: 10/23/2023
Certification File No.:	01734	Calib. No.: 00048
Linearity File No.:	01735	Cert. Date: 10/23/2023
Solution File No.:	01737	Cert. No.: 00043
Sequential File No.:	01737	Lin. Date: 10/23/2023
		Lin. No.: 00043
		Soln. Date: 10/28/2023
		Soln. No.: 00325
		File Date: 10/28/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	23110	Expires: 04/12/2025
		Bottle No.: 1316

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:36D		
Control 1 EC	0.100%	14:37D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	14:37D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:38D		
Control 2 EC	0.099%	14:38D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:38D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:39D		
Control 3 EC	0.099%	14:40D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:41D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_



#1123

Badge No.: 1123

Date: 10/28/2023