

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01733	Calib. Date: 10/23/2023
Certification File No.:	01734	Calib. No.: 00048
Linearity File No.:	01735	Cert. Date: 10/23/2023
Solution File No.:	01740	Cert. No.: 00043
Sequential File No.:	01740	Lin. Date: 10/23/2023
		Lin. No.: 00043
		Soln. Date: 11/29/2023
		Soln. No.: 00326
		File Date: 11/29/2023

Calibrating Unit:	WET	Model No.: CU-34	Serial No.: DDUL S3-0183
Control Solution %:	0.100%		Expires: 04/12/2025
Solution Control Lot:	23110		Bottle No.: 0524

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	09:34S		
Control 1 EC	0.096%	09:35S	33.8°C	*** TEST PASSED ***
Control 1 IR	0.097%	09:35S	33.8°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:42S		
Control 2 EC	0.097%	09:42S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	09:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:43S		
Control 3 EC	0.098%	09:44S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	09:44S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER

First Name: BENJAMIN

MI: L

Signature: _____



Badge No.: 1113

Date: 11/29/2023