

Calibrating Unit

New Standard Solution Report


Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01687	Calib. Date: 01/13/2023
Certification File No.:	01688	Cert. No.: 00046
Linearity File No.:	01689	Cert. Date: 01/13/2023
Solution File No.:	01693	Lin. No.: 00041
Sequential File No.:	01693	Lin. Date: 01/13/2023
		Soln. No.: 00314
		Soln. Date: 02/12/2023
		File Date: 02/12/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	22160	Expires: 05/09/2024
		Bottle No.: 1183

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:44S		
Control 1 EC	0.102%	18:45S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:46S		
Control 2 EC	0.098%	18:47S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:47S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:48S		
Control 3 EC	0.099%	18:48S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:48S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:49S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MILLER	First Name: BENJAMIN	MI: L
Signature: 		Badge No.: 1113
		Date: 02/12/2023