

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01687	Calib. Date: 01/13/2023
Certification File No.:	01688	Calib. No.: 00046
Linearity File No.:	01689	Cert. Date: 01/13/2023
Solution File No.:	01700	Cert. No.: 00041
Sequential File No.:	01700	Lin. Date: 01/13/2023
		Lin. No.: 00041
		Soln. Date: 05/09/2023
		Soln. No.: 00317
		File Date: 05/09/2023

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUL S3-0183
Control Solution %: 0.100%		Expires: 02/01/2025
Solution Control Lot: 23040		Bottle No.: 1338

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:15D		
Control 1 EC	0.104%	19:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	19:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:16D		
Control 2 EC	0.100%	19:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:18D		
Control 3 EC	0.100%	19:19D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	19:19D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:20D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_

Badge No.: 1123

Date: 05/09/2023