

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01701	Calib. Date: 06/06/2023
Certification File No.:	01702	Calib. No.: 00047
Linearity File No.:	01703	Cert. Date: 06/06/2023
Solution File No.:	01704	Cert. No.: 00042
Sequential File No.:	01704	Lin. Date: 06/06/2023
		Lin. No.: 00042
		Soln. Date: 06/06/2023
		Soln. No.: 00318
		File Date: 06/06/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 0720

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:43D		
Control 1 EC	0.102%	11:44D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:45D		
Control 2 EC	0.100%	11:46D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	11:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		
Control 3 EC	0.100%	11:47D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:48D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDUJ P2-035

Changed By:

Last Name: GUARCELLO

First Name: COREY

MI: R.

Signature:

Badge No.: 7035

Date: 06/06/2023