

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARTN-0042  
Location: HIGHTSTOWN POLICE  
Calibration File No.: 01701 Calib. Date: 06/06/2023 Calib. No.: 00047  
Certification File No.: 01702 Cert. Date: 06/06/2023 Cert. No.: 00042  
Linearity File No.: 01703 Lin. Date: 06/06/2023 Lin. No.: 00042  
Solution File No.: 01709 Soln. Date: 07/04/2023 Soln. No.: 00319  
Sequential File No.: 01709 File Date: 07/04/2023

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUL S3-0182  
Control Solution %: 0.100% Expires: 02/01/2025  
Solution Control Lot: 23040 Bottle No.: 1126

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:58D		
Control 1 EC	0.103%	13:59D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:59D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:59D		
Control 2 EC	0.101%	14:00D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	14:00D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:01D		
Control 3 EC	0.101%	14:02D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	14:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:02D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: \_\_\_\_\_



Badge No.: 1123

Date: 07/04/2023