

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01701	Calib. Date: 06/06/2023
Certification File No.:	01702	Calib. No.: 00047
Linearity File No.:	01703	Cert. Date: 06/06/2023
Solution File No.:	01721	Cert. No.: 00042
Sequential File No.:	01721	Lin. Date: 06/06/2023
		Lin. No.: 00042
		Soln. Date: 09/02/2023
		Soln. No.: 00321
		File Date: 09/02/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 1253

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	13:53D		
Control 1 EC	-.---%	13:54D	34.0°C	CTRL GAS SUPPLY
Control 1 IR	-.---%	13:54D	34.0°C	CTRL GAS SUPPLY
Ambient Air Blank	-.---%	--:--		
Control 2 EC	-.---%	--:--	N/A	
Control 2 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		
Control 3 EC	-.---%	--:--	N/A	
Control 3 IR	-.---%	--:--	N/A	
Ambient Air Blank	-.---%	--:--		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____



#1123

Badge No.: 1123

Date: 09/02/2023

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01701	Calib. Date: 06/06/2023
Certification File No.:	01702	Calib. No.: 00047
Linearity File No.:	01703	Cert. Date: 06/06/2023
Solution File No.:	01722	Cert. No.: 00042
Sequential File No.:	01722	Lin. Date: 06/06/2023
		Lin. No.: 00042
		Soln. Date: 09/02/2023
		Soln. No.: 00322
		File Date: 09/02/2023
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0182
Solution Control Lot:	23110	Expires: 04/12/2025
		Bottle No.: 1318

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:05D		
Control 1 EC	0.101%	16:06D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:06D		
Control 2 EC	0.100%	16:07D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	16:07D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:08D		
Control 3 EC	0.100%	16:09D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	16:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____

Badge No.: 1123

Date: 09/02/2023