

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTN-0042
Location:	HIGHTSTOWN POLICE	
Calibration File No.:	01687	Calib. Date: 01/13/2023
Certification File No.:	01688	Calib. No.: 00046
Linearity File No.:	01689	Cert. Date: 01/13/2023
Solution File No.:	01698	Cert. No.: 00041
Sequential File No.:	01698	Lin. Date: 01/13/2023
		Lin. No.: 00041
		Soln. Date: 04/11/2023
		Soln. No.: 00316
		File Date: 04/11/2023

Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUL S3-0183
Solution Control Lot:	23040	Expires: 02/01/2025
		Bottle No.: 1335

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:28D		
Control 1 EC	0.102%	18:29D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	18:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:30D		
Control 2 EC	0.099%	18:31D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	18:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:32D		
Control 3 EC	0.100%	18:33D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:33D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:34D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: STEPHENSEN

First Name: SGT

MI: R

Signature: _____



Badge No.: 1123

Date: 04/11/2023